PO Box 518 620 Applegate St. Philomath, OR 97370



Scattering Meter Calibration Sheet

11/11/2022
Wavelength: 412

S/N BB3-7905

Use the following equation to obtain "scaled" output values:

$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor } \times \text{ (Output - Dark Counts)}$					
Scale Factor for 412 nm	=	1.102E-05	(m ⁻¹ sr ⁻¹)/counts		
Output	=	4130	counts		
Dark counts	=	50	counts		
Instrument Resolution	=	1.0	counts 1.10E-05 (m ⁻¹ sr ⁻¹)		

Definitions:

- Scale Factor: Calibration scale factor, $\beta(\theta_c)$ /counts. Refer to User's Guide for derivation.
- **Output**: Measured signal output of the scattering meter.
- Dark Counts: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

PO Box 518 620 Applegate St. Philomath, OR 97370



Scattering Meter Calibration Sheet

11/11/2022
Wavelength: 532

S/N BB3-7905

Use the following equation to obtain "scaled" output values:

$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor } \times \text{ (Output - Dark Counts)}$						
Scale Factor for 532 nm	=	6.619E-0	<mark>)6</mark> (m⁻¹sr⁻¹)/c	ounts		
Output	=	413	30 counts			
Dark Counts	=	50 counts				
Instrument Resolution	=	1.0	counts	6.62E-06 (m ⁻¹ sr ⁻¹)		

Definitions:

- Scale Factor: Calibration scale factor, $\beta(\theta_c)$ /counts. Refer to User's Guide for derivation.
- **Output**: Measured signal output of the scattering meter.
- Dark Counts: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

PO Box 518 620 Applegate St. Philomath, OR 97370



Scattering Meter Calibration Sheet

11/11/2022	
Wavelength:676	

S/N#: BB3-7905

Use the following equation to obtain "scaled" output values:

$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor } \times \text{ (Output - Dark Counts)}$						
Scale Factor for 676 nm	=	3.185E-06	(m ⁻¹ sr ⁻¹)/c	counts		
Output	=	4130	counts			
Dark counts	=	50	counts			
Instrument Resolution	=	1.0	counts	3.19E-06 (m ⁻¹ sr ⁻¹)		

Definitions:

- Scale Factor: Calibration scale factor, $\beta(\theta_c)$ /counts. Refer to User's Guide for derivation.
- Output: Measured signal output of the scattering meter.
- Dark Counts: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.